
Pomona®

Model 6726 Alligator Clip with 8/32 Threaded Jack



Features:

- ✓ Alligator clip screws onto probe tip, eliminating accidental test probe and alligator clip separation.
- ✓ Use with 8/32 threaded probes (e.g. Pomona model 6723)
- ✓ Complies to International Safety Standard IEC 1010-2-031

Materials:

Alligator contacts: Nickel Plated Steel
Threaded Insert: Nickel Plated Brass

Insulation: Nylon 6 Spring: Steel

Ratings:

Alligator Clip: 300 Vrms, CAT II, 10 A IEC 1010-2-031

Temperature: -10 C to +55 C

Ordering Information:

A set of red and black medium alligator test clips.

Fits Test Probe Kit:

Kit includes: A set of red and black test leads.

A set of red and black medium alligator test clips (e.g. 6726).

All dimensions are in inches. Tolerances (except noted): $.xx = \pm .02$ " (,51 mm), $.xxx = \pm .005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies. Made in USA

Sales: 800-490-2361 Fax: 888-403-3360 Technical Assistance: 800-241-2060 For "Where to Buy" information, visit the Pomona web site at www.pomonaelectronics.com

Pomona®

Model 6726 Alligator Clip with 8/32 Threaded Jack

^Domona[®]


The items included with this product are certified to meet IEC 1010-2-031 safety requirements.

Les élements inclus avec ce produit sont certifies pour répondre à des exigences de surete du IEC 1010-2-031. Los items incluidos con este producto se certifican para resolver requisitos de seguridad del IEC 1010-2-031.

 Only IEC 1010 type shrouded test plugs should be used with IEC 1010 type jacks. NEVER INSERT A PROBE TIP WITH BANANA SPRING INTO A TEST ACCESSORY JACK THAT WILL BE USED AS A HAND HELD PROBING DEVICE -Seuls les vérificateurs de tension renforcés de type IEC 1010 devraient être utilisés avec des jacks de type IEC 1010. NE JAMAIS INSERER DE BOUT DE SONDE AVEC RESSORT BANANE DANS UN JACK D'ESSAI D'ACCESSOIRES QUI SERA UTILISE COMME DISPOSITIF DE SONDAGE A MAIN.

-Se deben usar sólo tapones de prueba reforzados del tipo IEC 1010 con receptáculos de tipo IEC 1010. NUNCA INTRODUZCA UNA CLAVIJA CON UN MUELLE TIPO BANANA EN UN RECEPTACULO DE PRUEBA QUE A SU VEZ SE USE COMO APARATO DE PRUEBA MANUAL.





Overvoltage Category I, II, III or IV. CAT Surtension Catégorie I, II, III ou IV. Sobretensión de Categoria I, II III o IV.



Caution! Refer to Accompanying Documents. Attention! Se Reporter aux Documentsjoints. ¡Peligro! Consulte los Documentos Adjuntos.



Caution! Risk of Electric Shock. Attention! Risque de Choc Électrique. ¡Peligro! Riesgo de Sacudida Electrica.



All dimensions are in inches. Tolerances (except noted): $.xx = \pm .02$ " (,51 mm), $.xxx = \pm .005$ " (,127 mm). All specifications are to the latest revisions. Specifications are subject to change without notice. Registered trademarks are the property of their respective companies. Made in USA