Termination GS

A036

Termination GS 38 Compliant Widely used and accepted throughout the Industry



Voltage / Amperage Rating

Maximum Voltage: 1000V Maximum Amperage: 30mA

Copyright © 2007 Test Products International, Inc.

Agency Listing

cULus 61010 IEC 61010

Category Rating

CAT III 1000V

Insulation Material

N/A

Dimensions

Lead Length: N/A Prod Length: 0.12" (3.0mm) Jaw Opening: N/A

Type

0.16" (4.0mm) male banana plug